

THE UNIVERSITY OF WESTERN ONTARIO OCCUPATIONAL HEALTH AND SAFETY

CONTAINMENT LEVEL 2 - LABORATORY BIOSAFETY PRACTICES

For laboratories using infectious agents listed as requiring Level 2 precautions as listed in Laboratory

*16. Procedures generating aerosols must be carried out in a biological safety cabinet. Centrifugation must be carried out in closed containers.

