Western University
Nuclear Radiation Safety Inspection Checklist

5.6(d) Daily usage, remaining quantities and final disposal dates are recorded on the inventory/waste forms for all unsealed nuclear substances. 5.7 Licence to service radiation device: A company or person who performs installation, repair or dismantling a device other than routine operating procedures as indicated in the P D Q X I D F W X U H U ¶ V R S H U D W L Q J P D Q X D O I R U W K H G H Y L F 5.8 Post Permit: Radiation permit is posted in all designated radiation rooms with current information (i.e. workers, lab, etc.). 5.9 Record Keeping/Retained: Inventory/disposal records and measurements are retained for the minimum of three years.		
6. Organization and Management 6.1 : R U N H t¶V 3 Operations: Workers are working safely and ensuring security. 6.2 Area Classification: 6.2(a) All locations being used for handling or storing nuclear substance are authorized by the Radiation Safety Coordinator and are listed on the permit 6.2(b) All nuclear substances in storage and in use is within order/possession limits and indicated on the permit. Activity of nuclear substances handled on bench/fume hood are within room classification limits. 6.3 Supervision: All radiation users are listed on the permit. Permit holder or designate is available to supervise authorized radiation users and designated radiation rooms. 6.4 Change Notified: Permit holder reports any changes on the permit to the Radiation Safety Coordinator. 6.5 Radiation Safety Manual: Radiation safety manual is accessible to all workers listed on the radiation permit. 6.6 Prohibition of Human Use: No nuclear substances are being used in or on human beings.		